

A New High-Flexion Design Total Knee Prosthesis: A Nine to Eleven-Year Follow-Up Study.



Kim MS, Choi KY, Kwak DH, Cho RK, Jang HJ,
Yang SC, In Y

*Seoul St. Mary's Hospital
The Catholic University of Korea*



Conflict of interest

Kim MS, MD. Ph.D.

Choi KY, MD.

Kwak DH, MD.

Cho RK, MD.

Jang HJ, MD.

Yang SC, MD.

In Y, MD, Ph.D.

We have no financial conflict to disclose.

Introduction

- ◆ Does high-flexion TKA design improve postoperative flexion & functional outcome?
- ◆ High flexion TKA vs Standard TKA

1. *ROM*

- High flexion TKA > Standard TKA
- High flexion TKA = Standard TKA

Bin 2007, Gupta 2006, Huang 2005
Seon 2005, Kim 2005, McCalden 2009

2. *Early aseptic loosening*

- High flexion TKA > Standard TKA

Cho 2011, Han 2007

Purpose

During a 9-11 year follow-up

- (1) To evaluate the **postoperative ROM** of high-flexion concept TKA (LOSPA)
- (2) To evaluate the **clinical outcomes** of LOSPA prosthesis
- (3) To evaluate **stability including survival rate** of LOSPA prosthesis

Materials and Methods

◆ Preoperative & postoperative ROM

- ROM, Flexion contracture, Maximum flexion

◆ Preoperative & postoperative clinical outcomes

- WOMAC score (Pain, Stiffness, function)

- Validated, 24-item disease-specific questionnaire

◆ Radiographic assessment



- Knee alignment, Femoral & tibial angle, Radiolucent line

Materials and Methods

➤ Radiolucent line

- Loosening: ≥ 2 mm in width at the cement-bone or cement-prosthesis interface
- Presence and location of radiolucent lines

Gonzalez 2004



Ewald 1989

Results (ROM & Clinical outcome)

	Preoperative	Postoperative (Last f/u)	p value
ROM(°)	117.3 ± 12.4	126.5 ± 9.7	<0.001
Flexion contracture(°)	6.3 ± 6.2	0.3 ± 1.7	<0.001
Maximum flexion(°)	123.6 ± 10.8	126.8 ± 8.6	0.001

	Preoperative	Postoperative (Last f/u)	p value
WOMAC	56.3 ± 14.9	15.8 ± 14.1	<0.001
Pain	11.4 ± 3.4	1.5 ± 2.7	<0.001
Stiffness	3.7 ± 2.0	1.3 ± 1.5	<0.001
Function	41.0 ± 11.6	13.1 ± 11.4	<0.001

Results (Radiologic findings)

	Preoperative	Postoperative (Last f/u)	<i>p</i> value
Knee alignment			
HKA(°)	Varus 7.5 ± 5.2 (Varus 25.2 – valgus 7.0)	Varus 1.5 ± 2.6 (Varus 14.0 – valgus 9.0)	<0.001
FTA(°)	Varus 3.5 ± 5.4 (Varus 19.8 – valgus 12.3)	Valgus 5.2 ± 2.4 (Varus 4.7 – valgus 18.8)	<0.001

Femoral angle		Tibial angle	
Coronal(°)	Sagittal(°)	Coronal(°)	Sagittal(°)
96.3 ± 2.0 (85.8 – 107.3)	4.9 ± 3.1 (0 – 14.3)	89.1 ± 1.8 (81.6 – 97.8)	86.2 ± 2.1 (79.2 – 96.0)

Results(Radiolucent lines)

Radiolucent line	23 knees (4.6%)
-------------------------	-----------------

Location	N
-----------------	----------

Medial tibial condyle (zone 1 & 2)	10
---	----

Anterior femoral condyle (zone 1 & 2)	12
--	----

Posterior femoral condyle (zone 3 & 4)	5
---	---

Four knees with radiolucent lines in both the anterior & posterior femoral condyle

Results(Survival rates)

Aseptic loosening,

Recurvatum, Ankylosis

+Infection

+PE change

+Brisement

+periprosthetic fx

Aseptic loosening,

Recurvatum, Ankylosis

+ Infection

Aseptic loosening,

Recurvatum, Ankylosis

Survival
rates (%)

96.2 %

97.2 %

97.8 %

Conclusion

- ◆ *High-flexion concept LOSPA system*

- (1) Improved ROM & clinical outcomes
- (2) Good survival rate after long-term follow up

Reference

1. Heiberg K, Bruun-Olsen V, Mengshoel AM. Pain and recovery of physical functioning nine months after total knee arthroplasty. *J Rehabil Med.* 2010;42:614–9.
2. Heck DA, Robinson RL, Partridge CM, Lubitz RM, Freund DA. Patient outcomes after knee replacement. *Clin Orthop Relat Res.* 1998;93–110.
3. Kurosaka M, Yoshiya S, Mizuno K, Yamamoto T. Maximizing flexion after total knee arthroplasty: the need and the pitfalls. *J Arthroplasty.* 2002;17:59–62.
4. Mulholland SJ, Wyss UP. Activities of daily living in non-Western cultures: range of motion requirements for hip and knee joint implants. *Int J Rehabil Res.* 2001;24:191–8.
5. Gupta SK, Ranawat AS, Shah V, Zikria BA, Zikria JF, Ranawat CS. The P.F.C. sigma RP-F TKA designed for improved performance: a matched-pair study. *Orthopedics.* 2006;29:S49–52.
6. Jones RE. High-flexion rotating-platform knees: rationale, design, and patient selection. *Orthopedics.* 2006;29:S76–9.
7. Bellemans J, Banks S, Victor J, Vandenneucker H, Moemans A. Fluoroscopic analysis of the kinematics of deep flexion in total knee arthroplasty. Influence of posterior condylar offset. *J Bone Joint Surg Br.* 2002;84:50–3.
8. Kim YH, Sohn KS, Kim JS. Range of motion of standard and high-flexion posterior stabilized total knee prostheses. A prospective, randomized study. *J Bone Joint Surg Am.* 2005;87:1470–5.
9. Argenson JN, Scuderi GR, Komistek RD, Scott WN, Kelly MA, Aubaniac JM. In vivo kinematic evaluation and design considerations related to high flexion in total knee arthroplasty. *J Biomech.* 2005;38:277–84.
10. Jain S, Pathak AC, Kannian K, Kulkarni S, Tawar S, Mane P. High-flexion posterior-stabilized total knee prosthesis: is it worth the hype? *Knee Surg Relat Res.* 2013;25:100–5.