



ISAKOS  
CONGRESS  
2025



MUNICH  
GERMANY  
June 8-11

# Catastrophic Full Thickness Wear of Oxinium Knee Implants: A Case Report and Analysis of Failure Mechanisms

1. Rajesh Malhotra, MS  
Orthopaedics, FIMSA, FRCS,  
FACS\*
2. Naman Wahal, MBBS, MS  
Orthopaedics\*

\*Indraprastha Apollo Hospital, New Delhi, India



# Faculty Disclosure Information

- Nothing to disclose



ISAKOS  
CONGRESS  
2025



**MUNICH**  
**GERMANY**  
June 8-11



# Background

- Oxinium (oxidized zirconium) implants
  - For patients with metal hypersensitivity
  - Improved wear resistance in knee arthroplasty
- Multiple reports of catastrophic wear
- Mechanism of failure not clear





# Case Report – History and Presentation

- 63-year-old female,
- 9 years post bilateral knee arthroplasty with oxinium implants.
- Symptoms: Progressive right knee pain and instability starting 2 years post-surgery.
- Radiographs: Medial joint space narrowing and metal debris suggestive of metallosis



**ISAKOS**  
CONGRESS  
2025



**MUNICH**  
**GERMANY**  
June 8-11



# Sequential Xrays



2015



2017



2023



2024



ISAKOS  
CONGRESS  
2025



MUNICH  
GERMANY  
June 8-11



# Intraoperative Findings

- Complete polyethylene wear on postero-medial aspect.
- Scratched oxinium surface.
- Black metallosis throughout joint.







# Mechanism of Failure

## 1. Polyethylene Wear:

- Led to metal-on-metal articulation

## 2. Edge Loading:

- Scratched the oxinium surface
- Exposed softer zirconium core

## 3. Metallosis:

- Triggered inflammation, causing synovitis and bone loss

## 4. Surgical Technique and Implant Design:

- Possible improper polyethylene seating.
- Unbalanced gaps contributed to early failure



**ISAKOS**  
CONGRESS  
2025



**MUNICH**  
**GERMANY**  
June 8-11

# Revision Surgery and Outcome

- LCCK-like implant used with distal and posterior femoral augments
- All visible metallosis debrided except posterior recess
- Patient allowed full weight bearing from day 1
- At 6 months: 0-70° ROM, no pain or instability
- At 1 year: 0-80° active ROM, 0-110 ° passive ROM, no pain or instability



# Post operative Xrays



Immediate  
Post-op



6 month  
follow-up



1 year follow-  
up



# Clinical picture at 6 months







# Lessons and Recommendations

- **Early detection and polyethylene insert exchange may prevent catastrophic failure**
- Frequent radiographic follow-up is essential
- Oxinium implant patients should be monitored closely, especially after trauma
- Surgical handling of implants must be meticulous



**ISAKOS**  
CONGRESS  
2025



**MUNICH**  
**GERMANY**  
June 8-11





# Summary and Conclusion

- Oxinium offers good wear resistance but can catastrophically fail when zirconium is exposed
- This case illustrates the cascade from polyethylene wear to metallosis
- Preventive strategies and better implant design are needed



# References

- **Frye BM**, Laughery KR, Klein AE. *The Oxinium Arthrogram: A Sign of Oxidized Zirconium Implant Failure*. Arthroplasty Today. 2021;8:103–106. [PMCID: PMC7943965]
- **Tribe H**, Malek S, Stammers J, Ranawat V, Skinner JA. *Advanced wear of an Oxinium™ femoral head implant following polyethylene liner dislocation*. Ann R Coll Surg Engl. 2013;95(8):e138–e141. [PMID: 24165329]
- **Greco NJ**, Berend KR. *Polyethylene liner dislocation of fixed-bearing medial Oxinium unicompartmental arthroplasty with severe metallosis*. Knee. 2018;25(2):341–345. [PMID: 29525546]
- **Kore L**, Bates T, Mills G, Lybeck D. *Oxidized Zirconium Total Knee Arthroplasty Implant Failure in a Patient With Knee Instability*. Arthroplasty Today. 2020;6(3):552–556. [PMCID: PMC7397700]
- **Purcell A**, Buckner S, Brindley G, Grimes J. *A unique case of extra-articular extravasation of metal into the lower leg resulting from oxidized zirconium wear particles from total knee arthroplasty*. Arthroplasty Today. 2020;6(4):988–992. [PMID: 33385039]
- **Cañizares S**, Barriga GC, Jarrín FV, Freire CDP. *Metallosis after Oxinium total knee arthroplasty in a patient with rheumatoid arthritis: A case report*. Cureus. 2023;15(2):e34759. [PMCID: PMC9985308]
- **Kelly B**, Manes T, Main C. *Bilateral failure of oxidized zirconium implants in total knee arthroplasty*. Arthroplasty Today. 2023;21:101098. [PMCID: PMC10192679]
- **Seon JK**, Ayob KA, Noh MG, Yang HY. *Peculiar reaction of oxidized zirconium from a total knee arthroplasty prosthesis: A case report*. Acta Orthop Traumatol Turc. 2024;58(1):68–72. [PMCID: PMC11059476]



**ISAKOS**  
CONGRESS  
2025



**MUNICH**  
**GERMANY**  
June 8–11